



Accredited Laboratory

A2LA has accredited

COSMO SOLUTIONS TECHNOLOGY, INC. - CST Calibration Laboratory

Farmington Hills, MI

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 10th day of June 2024.

A handwritten signature in blue ink, appearing to read 'Trace McInturff'.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 6929.01
Valid to July 31, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017



COSMO SOLUTIONS TECHNOLOGY, INC. - CST CALIBRATION LABORATORY
23855 Research Dr, Suite A
Farmington Hills, MI 48335
LabDirector@CosmoCalLab.org Phone: 248 488 2580

CALIBRATION

Valid To: July 31, 2026

Certificate Number: 6929.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations^{1,4}:

I. Fluid Quantities

Parameter/Equipment	Range	CMC ² , (±)	Comments
Gas Flow Rate	1 mL/min to 200 L/min	1.5 % rdg	Laminar Flow Measurement Reference standard
	1 mL/min to 200 L/min	1.8 % rdg	Laminar Flow Measurement Working standard
Gas Flow Rate ³	1 mL/min to 200 L/min	2.0 % rdg	Laminar Flow Measurement Working standard

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

COPY

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ This scope meets A2LA's *P112 Flexible Scope Policy*.